KEY LEARNING OBJECTIVES

- To determine whether surgery is indicated for an antebrachial deformity
- Have knowledge of the treatment options available for an antebrachial deformity
- Select appropriate surgical methods for presenting angular limb deformity cases

MULTIPLE CHOICE QUESTIONS

1. What percentage of ulna length does the distal ulna growth plate contribute?

(A) 100%	(C) 60%
(B) 80%	(D) 40%

2. In a dog that is under 5 months of age, over what valgus deformity would ulnar ostectomy be frequently unsuccessful?

(A) 10 degrees
(B) 15 degrees
(C) 20 degrees
(D) 25 degrees
3. What proportion of condrodystrophic dogs with antebrachial deformities have biapical deformities?

(A) 20%
(B) 40%
(C) 60%
(D) 80%

Getting the most out of surgical staplers

Rachel Hattersley

Surgical staplers are used in all aspects of veterinary surgery from skin closure to gastrointestinal surgery. Devices are usually tailored to a specific purpose and for each type of staple. Staple height is determined by the leg length of the staple after closure and the thickness of the tissue should be considered when deciding on which staple size to choose. There are five main types of staplers.

- Skin staplers. These are available in regular and wide dimensions. Staple height or leg length is also variable and is approximately 3.4–3.9 mm for regular staples and 3.9–4.1 mm for wide staples. Skin staplers tend to evert the skin edges and thus need to be placed carefully. The main advantage is speed of placement. The use of skin staplers has also been reported for closure of enterotomy wounds
- Vascular clips. Haemostatic clips can be useful for ligating vessels in challenging areas where access precludes the use of traditional suture ligatures. Reusable handpieces and disposable versions are available. The ligate and divide stapler (LDS) combines haemostatic clips with a carbon dioxide-powered blade which applies two clips and divides the tissue between the two. Such devices are single use only
- Thoracoabdominal (TA) stapler. Linear stapling devices are available in a variety of lengths (30–100 mm) and staple height (the colour of the cartridge usually denotes the staple height). These staplers fire either two (green or blue cartridges) or

three (vascular stapler which is white) staggered rows of B-shaped staples. These are predominantly used for liver and lung lobectomy but are also used in stapled gastrointestinal anastomosis. The handpieces are designed to be single-use which are reloadable for the same patient within the theatre. In reality, handpieces are often used more than once, but it should be noted this is not the manufacturer's recommendation

- Gastrointestinal (GIA) stapler. GIA staplers are also linear staplers which are combined with a cutting device. Two lines of staples are deployed on either side of a blade, which cuts the tissue after or while staples are placed. They are most commonly used for side-to-side gastrointestinal anastomosis but can also be used for cholecystoduodenostomy. Endoscopic linear staplers (endo-GIA) have staples that continue around the distal end of the blade to ensure that incomplete incision into vessels or hollow organs will not result in leakage
- Circular stapler. Circular stapling devices are designed to perform an end-to-end anastomosis in the gastrointestinal tract via an intraluminal approach. A double, staggered, circular row of staples is placed, and a stoma is created within the staples. The resulting anastomosis is inverted. Use of these devices is limited by the diameter of the bowel and they are therefore not commonly used in small animal surgery.

KEY LEARNING OBJECTIVES

- Become familiar with what surgical staplers are available
- Understand in what clinical situations to use surgical staplers
- Learn about complications of surgical staplers

Delivered by BSAVA to:

Guest (guest)

IP: 54.162.127.20

On: Wed, 26 Apr 2023 13:16:21

MULTIPLE CHOICE QUESTIONS

1. Which colour thoracoabdominal staple cartridge contains three rows of staggered B-shaped staples?
(A) White
(B) Blue
(C) Green
(D) Yellow
2. What is the height of the closed staple when a blue cartridge is used?
(A) 1.0 mm
(B) 1.5 mm
(C) 2.0 mm
(D) 2.5 mm
3. What kind of anastomosis is usually created by use of a linear GIA stapler when used in the small intestine?
(A) End-to-end
(B) Side-to-side
(C) Side-to-end
(D) End-to-side